



Certificate of Analysis

COMPLIANCE FOR RETAIL

Laboratory Sample ID: DA50422014-001



Apr 25, 2025 | Sunnyside

 22205 Sw Martin Hwy
 indiantown, FL, 34956, US


PASSED

Pages 1 of 1

SAFETY RESULTS


 Pesticides
PASSED

 Heavy Metals
PASSED

 Microbials
PASSED

 Mycotoxins
PASSED

 Residuals
 Solvents
PASSED

 Filtration
PASSED

 Water Activity
PASSED

 Moisture
NOT TESTED

 Terpenes
NOT TESTED

MISC.



Cannabinoid

TESTED

Total THC
0.245%

Total THC/Container : 102.657 mg


Total CBD
ND

Total CBD/Container : 0.000 mg


Total Cannabinoids
0.250%

Total Cannabinoids/Container : 104.752 mg

	D9-THC	THCA	CBD	CBDA	D8-THC	CBG	CBGA	CBN	THCV	CBDV	CBC
%	0.245	ND	ND	ND	ND	0.003	ND	ND	ND	ND	0.002
mg/unit	102.66	ND	ND	ND	ND	1.26	ND	ND	ND	ND	0.84
LOD	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
%	%	%	%	%	%	%	%	%	%	%	%

 Analyzed by:
 3335, 1665, 585, 1440

 Weight:
 3.0902g

 Extraction date:
 04/23/25 11:13:49

 Extracted by:
 3335

Analysis Method : SOP.T.40.031, SOP.T.30.031

Analytical Batch : DA085687POT

Instrument Used : DA-LC-007

Analyzed Date : 04/24/25 08:56:56

Batch Date : 04/23/25 08:03:43

Dilution : 40

Reagent : 030125.01; 012125.05; 021125.07; 041825.R05; 040725.R02

Consumables : 4052-629; 947.110; 040724CH01; 0000355309

Pipette : DA-108; GA-179; GA-178

Full Spectrum cannabinoid analysis utilizing High Performance Liquid Chromatography with UV detection in accordance with F.S. Rule 64ER20-39.

Label Claim

PASSED

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino

Lab Director

 State License # CMTL-0002
 ISO 17025 Accreditation # ISO/IEC
 17025:2017 Accreditation PJA-
 Testing 97164



 Signature
 04/25/25