



Certificate of Analysis

COMPLIANCE FOR RETAIL

Laboratory Sample ID: DA50430005-010



Production Method: Other - Not Listed

Harvest/Lot ID: 6467930630835261

Batch#: 6467930630835261

Cultivation Facility: FL - Indiantown (4430)

Processing Facility: FL - Indiantown (4430)

Source Facility: FL - Indiantown (4430)

Seed to Sale#: 1886064425944616

Harvest Date: 04/28/25

Sample Size Received: 16 units

Total Amount: 2806 units

Retail Product Size: 1 gram

Retail Serving Size: 1 gram

Servings: 1

Ordered: 04/30/25

Sampled: 04/30/25

Completed: 05/03/25

Sampling Method: SOP.T.20.010

May 03, 2025 | Sunnyside

22205 Sw Martin Hwy
indiantown, FL, 34956, US

Sunnyside*

PASSED

Pages 1 of 6

SAFETY RESULTS



Pesticides
PASSED



Heavy Metals
PASSED



Microbials
PASSED



Mycotoxins
PASSED



Residuals Solvents
PASSED



Filtration
PASSED



Water Activity
PASSED



Moisture
NOT TESTED



Terpenes
TESTED

MISC.



Cannabinoid

TESTED



Total THC
82.703%

Total THC/Container : 827.030 mg



Total CBD
0.137%

Total CBD/Container : 1.370 mg



Total Cannabinoids
92.987%

Total Cannabinoids/Container : 929.870 mg

	D9-THC	THCA	CBD	CBDa	D8-THC	CBG	CBGA	CBN	THCV	CBDV	CBC
%	14.546	77.717	ND	0.157	ND	0.166	0.207	ND	ND	ND	0.155
mg/unit	145.46	777.17	ND	1.57	ND	1.66	2.07	ND	ND	ND	1.55
LOD	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
%											

Analyzed by:
4351, 1665, 585, 1440

Weight:
0.1085g

Extraction date:
05/01/25 13:11:35

Extracted by:
4351

Analysis Method : SOP.T.40.031, SOP.T.30.031

Analytical Batch : DA085978POT

Instrument Used : DA-LC-003

Analyzed Date : 05/03/25 03:21:12

Batch Date : 05/01/25 08:54:44

Dilution : 400

Reagent : 021125.07; 042325.R30; 043025.R35

Consumables : 947.110; 04312111; 040724CH01; 1009429049; 1009372593; 0000355309

Pipette : DA-055; DA-063; DA-067

Full Spectrum cannabinoid analysis utilizing High Performance Liquid Chromatography with UV detection in accordance with F.S. Rule 64ER20-39.

Label Claim

PASSED

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino

Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJLA-
Testing 97164



Signature
05/03/25



Certificate of Analysis

PASSED

Sunnyside

22205 Sw Martin Hwy
Indiantown, FL, 34956, US
Telephone: (772) 631-0257
Email: Julio.Chavez@crescolabs.com

Sample : DA50430005-010
Harvest/Lot ID: 6467930630835261

Batch# : 6467930630835261 Sample Size Received : 16 units
Sampled : 04/30/25 Total Amount : 2806 units
Ordered : 04/30/25 Completed : 05/03/25 Expires: 05/03/26
Sample Method : SOP.T.20.010

Page 2 of 6

Terpenes					TESTED				
Terpenes	LOD (%)	Pass/Fail	mg/unit	Result (%)	Terpenes	LOD (%)	Pass/Fail	mg/unit	Result (%)
TOTAL TERPENES	0.007	TESTED	53.78	5.378	NEROL	0.007	TESTED	ND	ND
BETA-CARYOPHYLLENE	0.007	TESTED	12.31	1.231	PULEGONE	0.007	TESTED	ND	ND
LINALOOL	0.007	TESTED	9.20	0.920	SABINENE HYDRATE	0.007	TESTED	ND	ND
BETA-MYRCENE	0.007	TESTED	7.57	0.757	VALENCENE	0.007	TESTED	ND	ND
LIMONENE	0.007	TESTED	5.88	0.588	ALPHA-CEDRENE	0.005	TESTED	ND	ND
ALPHA-HUMULENE	0.007	TESTED	4.09	0.409	ALPHA-PHELLENDRENE	0.007	TESTED	ND	ND
FARNESENE	0.001	TESTED	3.23	0.323	ALPHA-TERPINENE	0.007	TESTED	ND	ND
FENCHYL ALCOHOL	0.007	TESTED	2.39	0.239	CIS-NEROLIDOL	0.003	TESTED	ND	ND
TRANS-NEROLIDOL	0.005	TESTED	2.30	0.230	Analysis Method : SOP.T.30.061A.FL, SOP.T.40.061A.FL	Weight: 0.2407g	Extraction date: 05/01/25 12:29:15	Extracted by: 4644	
ALPHA-TERPINEOL	0.007	TESTED	1.39	0.139	Analyzed by: 6444, 889, 5440				
BETA-PINENE	0.007	TESTED	1.01	0.101	Analytical Batch : DA085991ITER				Batch Date : 05/01/25 10:09:42
BORNEOL	0.013	TESTED	0.91	0.091	Instrument Used : DA-GCMS-004				
CARYOPHYLLENE OXIDE	0.007	TESTED	0.54	0.054	Analyzed Date : 05/03/25 13:34:13				
OCHRENE	0.007	TESTED	0.53	0.053	Dilution : 10				
ALPHA-PINENE	0.007	TESTED	0.51	0.051	Reagent : 02525.51				
ALPHA-BISBOLOL	0.007	TESTED	0.43	0.043	Consumables : 947.110, 04312111; 2240626; 0000355309				
ALPHA-TERPINOLENE	0.007	TESTED	0.39	0.039	Pipette : DA-065				
CAMPHERE	0.007	TESTED	0.32	0.032	Terpenoid testing is performed utilizing Gas Chromatography Mass Spectrometry. For all Flower samples, the Total Terpenes % is dry-weight corrected.				
FENCHONE	0.007	TESTED	0.30	0.030					
GAMMA-TERPINENE	0.007	TESTED	0.28	0.028					
SABINENE	0.007	TESTED	0.20	0.020					
3-CARENE	0.007	TESTED	ND	ND					
CAMPHOR	0.007	TESTED	ND	ND					
CEDROL	0.007	TESTED	ND	ND					
EUCALYPTOL	0.007	TESTED	ND	ND					
GERANOL	0.007	TESTED	ND	ND					
GERANYL ACETATE	0.007	TESTED	ND	ND					
GUAIOL	0.007	TESTED	ND	ND					
HEXAHYDROTHYMOL	0.007	TESTED	ND	ND					
ISOBORNEOL	0.007	TESTED	ND	ND					
ISOPULEGOL	0.007	TESTED	ND	ND					
Total (%)				5.378					

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino
Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation P/LA-
Testing 97164

Signature
05/03/25



Certificate of Analysis

PASSED

Sunnyside

Sample : DA50430005-010
Harvest/Lot ID: 6467930630835261

22205 Sw Martin Hwy
indiantown, FL, 34956, US
Telephone: (772) 631-0257
Email: Julio.Chavez@crescolabs.com

Batch# : 6467930630835261 Sample Size Received : 16 units
Sampled : 04/30/25 Total Amount : 2806 units
Ordered : 04/30/25 Completed : 05/03/25 Expires: 05/03/26
Sample Method : SOP.T.20.010

Page 3 of 6



Pesticides

PASSED

Pesticide	LOD	Units	Action Level	Pass/Fail	Result	Pesticide	LOD	Units	Action Level	Pass/Fail	Result
TOTAL CONTAMINANT LOAD (PESTICIDES)	0.010	ppm	5	PASS	ND	OXAMYL	0.010	ppm	0.5	PASS	ND
TOTAL DIMETHOMORPH	0.010	ppm	0.2	PASS	ND	PACLOBUTRAZOL	0.010	ppm	0.1	PASS	ND
TOTAL PERMETHRIN	0.010	ppm	0.1	PASS	ND	PHOSMET	0.010	ppm	0.1	PASS	ND
TOTAL PYRETHRINS	0.010	ppm	0.5	PASS	ND	PIPERONYL BUTOXIDE	0.010	ppm	3	PASS	ND
TOTAL SPINETORAM	0.010	ppm	0.2	PASS	ND	PRALLETHRIN	0.010	ppm	0.1	PASS	ND
TOTAL SPINOSAD	0.010	ppm	0.1	PASS	ND	PROPICONAZOLE	0.010	ppm	0.1	PASS	ND
ABAMECTIN B1A	0.010	ppm	0.1	PASS	ND	PROPOXUR	0.010	ppm	0.1	PASS	ND
ACEPHATE	0.010	ppm	0.1	PASS	ND	PYRIDABEN	0.010	ppm	0.2	PASS	ND
ACEQUINOCYL	0.010	ppm	0.1	PASS	ND	SPIROMESIFEN	0.010	ppm	0.1	PASS	ND
ACETAMIPRID	0.010	ppm	0.1	PASS	ND	SPIROTETRAMAT	0.010	ppm	0.1	PASS	ND
ALDICARB	0.010	ppm	0.1	PASS	ND	SPIROXAMINE	0.010	ppm	0.1	PASS	ND
AZOXYSTROBIN	0.010	ppm	0.1	PASS	ND	TEBUCONAZOLE	0.010	ppm	0.1	PASS	ND
BIFENAZATE	0.010	ppm	0.1	PASS	ND	THIACLOPRID	0.010	ppm	0.1	PASS	ND
BIFENTHRIN	0.010	ppm	0.1	PASS	ND	THIAMETHOXAM	0.010	ppm	0.5	PASS	ND
BOSCALID	0.010	ppm	0.1	PASS	ND	TRIFLOXYSTROBIN	0.010	ppm	0.1	PASS	ND
CARBARYL	0.010	ppm	0.5	PASS	ND	PENTACHLORONITROBENZENE (PCNB) *	0.010	ppm	0.15	PASS	ND
CARBOFURAN	0.010	ppm	0.1	PASS	ND	PARATHION-METHYL *	0.010	ppm	0.1	PASS	ND
CHLORANTRANILIPROLE	0.010	ppm	1	PASS	ND	CAPTAN *	0.070	ppm	0.7	PASS	ND
CHLORMEQUAT CHLORIDE	0.010	ppm	1	PASS	ND	CHLORDANE *	0.010	ppm	0.1	PASS	ND
CHLORPYRIFOS	0.010	ppm	0.1	PASS	ND	CHLORFENAPYR *	0.010	ppm	0.1	PASS	ND
CLOFENTEZINE	0.010	ppm	0.2	PASS	ND	CYFLUTHRIN *	0.050	ppm	0.5	PASS	ND
COUMAPHOS	0.010	ppm	0.1	PASS	ND	CYPERMETHRIN *	0.050	ppm	0.5	PASS	ND
DAMINOZIDE	0.010	ppm	0.1	PASS	ND						
DIAZINON	0.010	ppm	0.1	PASS	ND						
DICHLORVOS	0.010	ppm	0.1	PASS	ND						
DIMETHOATE	0.010	ppm	0.1	PASS	ND						
ETHOPROPHOS	0.010	ppm	0.1	PASS	ND						
ETOFENPROX	0.010	ppm	0.1	PASS	ND						
ETOXAZOLE	0.010	ppm	0.1	PASS	ND						
FENHEXAMID	0.010	ppm	0.1	PASS	ND						
FENOXYCARB	0.010	ppm	0.1	PASS	ND						
FENPYROXIMATE	0.010	ppm	0.1	PASS	ND						
FIPRONIL	0.010	ppm	0.1	PASS	ND						
FLONICAMID	0.010	ppm	0.1	PASS	ND						
FLUDIOXONIL	0.010	ppm	0.1	PASS	ND						
HEXYTHIAZOX	0.010	ppm	0.1	PASS	ND						
IMAZALIL	0.010	ppm	0.1	PASS	ND						
IMIDACLOPRID	0.010	ppm	0.4	PASS	ND						
KRESOXIM-METHYL	0.010	ppm	0.1	PASS	ND						
MALATHION	0.010	ppm	0.2	PASS	ND						
METALAXYL	0.010	ppm	0.1	PASS	ND						
METHIACARB	0.010	ppm	0.1	PASS	ND						
METHOMYL	0.010	ppm	0.1	PASS	ND						
MEVINPHOS	0.010	ppm	0.1	PASS	ND						
MYCLOBUTANIL	0.010	ppm	0.1	PASS	ND						
NALED	0.010	ppm	0.25	PASS	ND						

Analyzed by: 3621, 585, 1440 Weight: 0.2581g Extraction date: 05/01/25 12:19:03 Extracted by: 3621

Analysis Method : SOP.T.30.102.FL, SOP.T.40.102.FL

Analytical Batch : DA086001PES

Instrument Used : DA-LCMS-004 (PES)

Batch Date : 05/01/25 10:30:42

Analyzed Date : 05/02/25 09:22:42

Dilution : 250

Reagent : 043025.R29; 043025.R28; 042925.R27; 042825.R02; 042925.R13; 043025.R04; 081023.01

Consumables : 6822423-02

Pipette : DA-093; DA-094; DA-219

Testing for agricultural agents is performed utilizing Liquid Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.

Analyzed by: 450, 585, 1440 Weight: 0.2581g Extraction date: 05/01/25 12:19:03 Extracted by: 3621

Analysis Method : SOP.T.30.151A.FL, SOP.T.40.151.FL

Analytical Batch : DA086003VOL

Instrument Used : DA-GCMS-011

Batch Date : 05/01/25 10:32:17

Analyzed Date : 05/02/25 09:19:28

Dilution : 250

Reagent : 042925.R27; 081023.01; 042325.R52; 042325.R53

Consumables : 6822423-02; 040724CH01; 17473601

Pipette : DA-080; DA-146; DA-218

Testing for agricultural agents is performed utilizing Gas Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.



Certificate of Analysis

PASSED

Sunnyside

 22205 Sw Martin Hwy
 indiantown, FL, 34956, US
 Telephone: (772) 631-0257
 Email: Julio.Chavez@crescolabs.com

 Sample : DA50430005-010
 Harvest/Lot ID: 6467930630835261

 Batch# : 6467930630835261 Sample Size Received : 16 units
 Sampled : 04/30/25 Total Amount : 2806 units
 Ordered : 04/30/25 Completed : 05/03/25 Expires: 05/03/26
 Sample Method : SOP.T.20.010

Page 4 of 6



Residual Solvents

PASSED

Solvents	LOD	Units	Action Level	Pass/Fail	Result
1,1-DICHLOROETHENE	0.800	ppm	8	PASS	ND
1,2-DICHLOROETHANE	0.200	ppm	2	PASS	ND
2-PROPANOL	50.000	ppm	500	PASS	ND
ACETONE	75.000	ppm	750	PASS	ND
ACETONITRILE	6.000	ppm	60	PASS	ND
BENZENE	0.100	ppm	1	PASS	ND
BUTANES (N-BUTANE)	500.000	ppm	5000	PASS	<2500.000
CHLOROFORM	0.200	ppm	2	PASS	ND
DICHLOROMETHANE	12.500	ppm	125	PASS	ND
ETHANOL	500.000	ppm	5000	PASS	ND
ETHYL ACETATE	40.000	ppm	400	PASS	ND
ETHYL ETHER	50.000	ppm	500	PASS	ND
ETHYLENE OXIDE	0.500	ppm	5	PASS	ND
HEPTANE	500.000	ppm	5000	PASS	ND
METHANOL	25.000	ppm	250	PASS	ND
N-HEXANE	25.000	ppm	250	PASS	ND
PENTANES (N-PENTANE)	75.000	ppm	750	PASS	ND
PROPANE	500.000	ppm	5000	PASS	ND
TOLUENE	15.000	ppm	150	PASS	ND
TOTAL XYLENES	15.000	ppm	150	PASS	ND
TRICHLOROETHYLENE	2.500	ppm	25	PASS	ND

Analyzed by: 4451, 585, 1440	Weight: 0.0202g	Extraction date: 05/01/25 11:48:35	Extracted by: 4451
---------------------------------	--------------------	---------------------------------------	-----------------------

 Analysis Method : SOP.T.40.041.FL
 Analytical Batch : DA08598750L
 Instrument Used : DA-GCMS-012
 Analyzed Date : 05/03/25 09:20:54

Batch Date : 05/01/25 09:55:03

 Dilution : 1
 Reagent : 030420.09
 Consumables : 429651; 315545
 Pipette : DA-309 25 uL Syringe 35028

Residual solvents analysis is performed utilizing Gas Chromatography Mass Spectrometry in accordance with with F.S. Rule 64ER20-39.



Certificate of Analysis

PASSED

Sunnyside

22205 Sw Martin Hwy
indiantown, FL, 34956, US
Telephone: (772) 631-0257
Email: Julio.Chavez@crescolabs.com

Sample : DA50430005-010
Harvest/Lot ID: 6467930630835261
Batch# : 6467930630835261 Sample Size Received : 16 units
Sampled : 04/30/25 Total Amount : 2806 units
Ordered : 04/30/25 Completed : 05/03/25 Expires: 05/03/26
Sample Method : SOP.T.20.010

Page 5 of 6

	Microbial	PASSED		Mycotoxins	PASSED
---	------------------	---------------	---	-------------------	---------------

Analyte	LOD	Units	Result	Pass / Fail	Action Level
ASPERGILLUS TERREUS			Not Present	PASS	
ASPERGILLUS NIGER			Not Present	PASS	
ASPERGILLUS FUMIGATUS			Not Present	PASS	
ASPERGILLUS FLAVUS			Not Present	PASS	
SALMONELLA SPECIFIC GENE			Not Present	PASS	
ECOLI SHIGELLA			Not Present	PASS	
TOTAL YEAST AND MOLD	10	CFU/g	<10	PASS	100000

Analyzed by: 4520, 585, 1440 **Weight:** 0.924g **Extraction date:** 05/01/25 09:33:22 **Extracted by:** 4520,4571
Analysis Method : SOP.T.40.056C, SOP.T.40.058.FL, SOP.T.40.209.FL
Analytical Batch : DA085976MIC
Instrument Used : PathogenDx Scanner DA-111,Applied Biosystems 2720 Thermocycler DA-010,Fisher Scientific Isotemp Heat Block (95°C) DA-049,DA-402 Thermo Scientific Heat Block (55 C) **Batch Date :** 05/01/25 08:26:12
Analyzed Date : 05/02/25 10:18:23
Dilution : 10
Reagent : 022625.43; 022625.58; 041525.R13; 080724.11
Consumables : 7582001005
Pipette : N/A

Analyte	LOD	Units	Result	Pass / Fail	Action Level
AFLATOXIN B2	0.002	ppm	ND	PASS	0.02
AFLATOXIN B1	0.002	ppm	ND	PASS	0.02
OCHRATOXIN A	0.002	ppm	ND	PASS	0.02
AFLATOXIN G1	0.002	ppm	ND	PASS	0.02
AFLATOXIN G2	0.002	ppm	ND	PASS	0.02

Analyzed by: 3621, 585, 1440 **Weight:** 0.2581g **Extraction date:** 05/01/25 12:19:03 **Extracted by:** 3621
Analysis Method : SOP.T.30.102.FL, SOP.T.40.102.FL
Analytical Batch : DA086002MYC
Instrument Used : N/A **Batch Date :** 05/01/25 10:32:15
Analyzed Date : 05/02/25 07:59:27
Dilution : 250
Reagent : 043025.R29; 043025.R28; 042925.R27; 042825.R02; 042925.R13; 043025.R04; 081023.01
Consumables : 6822423-02
Pipette : DA-093; DA-094; DA-219
 Mycotoxins testing utilizing Liquid Chromatography with Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.

Analyte	LOD	Units	Result	Pass / Fail	Action Level
TOTAL CONTAMINANT LOAD METALS	0.080	ppm	ND	PASS	1.1
ARSENIC	0.020	ppm	ND	PASS	0.2
CADMIUM	0.020	ppm	ND	PASS	0.2
MERCURIUM	0.020	ppm	ND	PASS	0.2
LEAD	0.020	ppm	ND	PASS	0.5

Analyzed by: 1022, 585, 1440 **Weight:** 0.2585g **Extraction date:** 05/01/25 12:08:48 **Extracted by:** 1022,4531
Analysis Method : SOP.T.30.082.FL, SOP.T.40.082.FL
Analytical Batch : DA085986HEA
Instrument Used : DA-ICPMS-004 **Batch Date :** 05/01/25 09:52:43
Analyzed Date : 05/02/25 10:14:22
Dilution : 50
Reagent : 041425.R05; 042225.R05; 042825.R05; 042125.R17; 042825.R03; 042825.R04; 120324.07; 042225.R04
Consumables : 040724CH01; J609879-0193; 179436
Pipette : DA-061; DA-191; DA-216
 Heavy Metals analysis is performed using Inductively Coupled Plasma Mass Spectrometry in accordance with F.S. Rule 64ER20-39.

Total yeast and mold testing is performed utilizing MPN and traditional culture based techniques in accordance with F.S. Rule 64ER20-39.

	Heavy Metals	PASSED
---	---------------------	---------------

Metal	LOD	Units	Result	Pass / Fail	Action Level
TOTAL CONTAMINANT LOAD METALS	0.080	ppm	ND	PASS	1.1
ARSENIC	0.020	ppm	ND	PASS	0.2
CADMIUM	0.020	ppm	ND	PASS	0.2
MERCURIUM	0.020	ppm	ND	PASS	0.2
LEAD	0.020	ppm	ND	PASS	0.5

Analyzed by: 1022, 585, 1440 **Weight:** 0.2585g **Extraction date:** 05/01/25 12:08:48 **Extracted by:** 1022,4531
Analysis Method : SOP.T.30.082.FL, SOP.T.40.082.FL
Analytical Batch : DA085986HEA
Instrument Used : DA-ICPMS-004 **Batch Date :** 05/01/25 09:52:43
Analyzed Date : 05/02/25 10:14:22
Dilution : 50
Reagent : 041425.R05; 042225.R05; 042825.R05; 042125.R17; 042825.R03; 042825.R04; 120324.07; 042225.R04
Consumables : 040724CH01; J609879-0193; 179436
Pipette : DA-061; DA-191; DA-216
 Heavy Metals analysis is performed using Inductively Coupled Plasma Mass Spectrometry in accordance with F.S. Rule 64ER20-39.



Certificate of Analysis

PASSED

Sunnyside

22205 Sw Martin Hwy
indiantown, FL, 34956, US
Telephone: (772) 631-0257
Email: Julio.Chavez@crescolabs.com

Sample : DA50430005-010
Harvest/Lot ID: 6467930630835261
Batch# : 6467930630835261 Sample Size Received : 16 units
Sampled : 04/30/25 Total Amount : 2806 units
Ordered : 04/30/25 Completed : 05/03/25 Expires: 05/03/26
Sample Method : SOP.T.20.010

Page 6 of 6

	Filth/Foreign Material	PASSED
---	-------------------------------	---------------

Analyte	LOD	Units	Result	P/F	Action Level
Filth and Foreign Material	0.100	%	ND	PASS	1

Analyzed by: 1879, 585, 1440	Weight: 1g	Extraction date: 05/01/25 12:04:07	Extracted by: 1879
---------------------------------	---------------	---------------------------------------	-----------------------

Analysis Method : SOP.T.40.090
Analytical Batch : DA086015FIL
Instrument Used : Filth/Foreign Material Microscope Batch Date : 05/01/25 11:53:53
Analyzed Date : 05/01/25 12:22:04

Dilution : N/A
Reagent : N/A
Consumables : N/A
Pipette : N/A

Filth and foreign material inspection is performed by visual inspection utilizing naked eye and microscope technologies in accordance with F.S. Rule 64ER20-39.

	Water Activity	PASSED
---	-----------------------	---------------

Analyte	LOD	Units	Result	P/F	Action Level
Water Activity	0.010	aw	0.506	PASS	0.85

Analyzed by: 4797, 585, 1440	Weight: 0.5534g	Extraction date: 05/01/25 13:47:23	Extracted by: 4797
---------------------------------	--------------------	---------------------------------------	-----------------------

Analysis Method : SOP.T.40.019
Analytical Batch : DA086006WAT
Instrument Used : DA-028 Rotronic HygroPalm Batch Date : 05/01/25 10:34:08
Analyzed Date : 05/02/25 07:46:30

Dilution : N/A
Reagent : 101724.36
Consumables : PS-14
Pipette : N/A

Water Activity is performed using a Rotronic HygroPalm HP 23-AW in accordance with F.S. Rule 64ER20-39.

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino
Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJLA-
Testing 97164



Signature
05/03/25